



CONSORTIUM of NRLs for KIT QUALITY

Under NACP

E-mail : consortium.nari@gmail.com
Website : www.nari-icmr.res.in

☎ : 020-27331200, 27331210
Fax No. : 0091-20-27121071

Ref: NARI/KQT/25/35

19/ 05/ 2025

Consortium of NRLs on Q

Batches evaluated under 'Consortium of NRLs for kit quality' that passed the criteria for sensitivity and specificity as per CDSCO guidelines:

HIV antibody detecting kits evaluated during January-March 2025:

Name of kit	Type of kit	Batch no.	Sensitivity	Specificity	Status
COMBAIDS®-RS Advantage-ST HIV 1+2	Rapid	4000031615	100 %	100%	Passed
		4000031735	100 %	100%	Passed
MI0325024		100%	100%	Passed	
MI0325027		100%	100%	Passed	

HIV-Syphilis duo antibody detecting kits evaluated during January-March 2025:

Name of kit	Type of kit	Batch no.	Sensitivity	Specificity	Status
Meriscreen HIV+ SYP Antibody Test	Rapid	MI1224084	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed
		MI0325054	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed
		MI0325055	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed
		MI1224086	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed
		MI1224087	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed
		MI1224088	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed
		MI0225065	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed
		MI0225070	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed
		MI0225071	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed

Consortium Members

ICMR-National Institute of Translational Virology and AIDS Research (NITYAR)
Plot No. 73, 'G' Block,
MIDC, Bhosari, Pune-411026.
Phone : 020-27331210
Fax : 0091-20-27121071
E-mail : consortium.nari@gmail.com

National Centre for Disease Control (NCDC)
22, Sham Nath Marg,
Delhi-110054
Phone: 011-23934517
E-mail :
consortiumncdc2@gmail.com

National Institute of Mental Health and Neuro Sciences (NIMHANS)
Hosur Road, Bengaluru-560029
Phone: 080-26995126, 26995778
Fax: 91-080-26564830, 6562121
Email :
nimhansnrlonq@gmail.com

ICMR-National Institute for Research in Bacterial Infections (NIRBI)
P-33, C.I.T. Road, Scheme XM,
Beliaghata, Kolkata-700010
Phone : +91-33-23633373/ 2370 1176
Fax : +91-(0)33-2370-5066
Email : niced.naco.nrlonq@gmail.com



CONSORTIUM of NRLs for KIT QUALITY

Under NACP

E-mail : consortium.nari@gmail.com
Website : www.nari-icmr.res.in

☎ : 020-27331200, 27331210
Fax No. : 0091-20-27121071

Name of kit	Type of kit	Batch no.	Sensitivity	Specificity	Status
Standard Q HIV/ Syphilis Combo Kit	Rapid	1CO3324002	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed
		1CO3325001	HIV-100 %, SYP- 88.75 %	HIV-100 %, SYP-99.5 %	Passed
		1CO3325002	HIV-100 %, SYP- 91.25 %	HIV-100 %, SYP-99.25 %	Passed
		1CO3325003	HIV-100 %, SYP- 90 %	HIV-100 %, SYP- 99.5 %	Passed
		1CO3325004	HIV-100 %, SYP- 91.25 %	HIV-100 %, SYP- 99.25 %	Passed
		1CO3325005	HIV-100 %, SYP- 92.5 %	HIV-100 %, SYP- 99.5 %	Passed
		1CO3325006	HIV-100 %, SYP- 91.25 %	HIV-100 %, SYP- 99.25 %	Passed
		1CO3325009	HIV-100 %, SYP- 100 %	HIV-100 %, SYP-100 %	Passed

Ashwini Shete
19.5.2025
Dr. Ashwini Shete
Scientist E, In-charge,
Consortium of NRL's for Kit Quality
ICMR-National Institute of
Translational Virology and AIDS
Research (ICMR-NITVAR)
Pune -411026

Sheela Godbole
Dr. Sheela Godbole
Director
ICMR-National Institute of Translational
Virology and AIDS Research (ICMR-
NITVAR)
Pune-411026

Consortium Members

ICMR-National Institute of Translational Virology and AIDS Research (NITVAR)
Plot No. 73, 'G' Block,
MIDC, Bhosari, Pune-411026.
Phone : 020-27331210
Fax : 0091-20-27121071
E-mail : consortium.nari@gmail.com

National Centre for Disease Control (NCDC)
22, Sham Nath Marg,
Delhi-110054
Phone: 011-23934517
E-mail :
consortiumncdc2@gmail.com

National Institute of Mental Health and Neuro Sciences (NIMHANS)
Hosur Road, Bengaluru-560029
Phone: 080-26995126, 26995778
Fax: 91-080-26564830, 6562121
Email :
nimhansnrlonq@gmail.com

ICMR-National Institute for Research in Bacterial Infections (NIRBI)
P-33, C.I.T. Road, Scheme XM,
Beliaghata, Kolkata-700010
Phone : +91-33-23633373/ 2370 1176
Fax : +91-(0)33-2370-5066
Email : niced.naco.nrlonq@gmail.com